

Fast shape-reconstruction in electrical impedance tomography

Bastian von Harrach

bastian.harrach@uni-wuerzburg.de

(joint work with Marcel Ullrich)

Institut für Mathematik - IX, Universität Würzburg

The Sixth International Conference
"Inverse Problems: Modeling and Simulation",
Antalya, Turkey, May 21–26, 2012.

Forward operator of EIT:

$$\Lambda : \sigma \mapsto \Lambda(\sigma), \quad \text{"conductivity"} \mapsto \text{"measurements"}$$

- ▶ Conductivity: $\sigma \in L_+^\infty(\Omega)$
- ▶ Continuum model: $\Lambda(\sigma)$: Neumann-Dirichlet-operator

$$\Lambda(\sigma) : g \mapsto u|_{\partial\Omega}, \quad \text{"applied current"} \mapsto \text{"measured voltage"}$$
$$\nabla \cdot (\sigma \nabla u) = 0 \quad \text{in } \Omega, \quad \sigma \partial_\nu u|_{\partial\Omega} = g \quad \text{on } \partial\Omega.$$

- ▶ Linear elliptic PDE theory:

$$\Lambda(\sigma) : L_\diamond^2(\partial\Omega) \rightarrow L_\diamond^2(\partial\Omega) \text{ linear, compact, self-adjoint}$$

Non-linear forward operator of EIT

$$\Lambda : \sigma \mapsto \Lambda(\sigma), \quad L_+^\infty(\Omega) \rightarrow \mathcal{L}(L_\diamond^2(\partial\Omega))$$

Inverse problem of EIT:

$$\Lambda(\sigma) \mapsto \sigma?$$

- ▶ Uniqueness ("Calderón problem"): Is Λ injective?
- ▶ Convergent numerical methods to reconstruct σ ?

Convergent numerical methods to reconstruct σ ?

- ▶ Newton iteration: almost no theory
Dobson (1992): (Local) convergence for regularized EIT equation.
Lechleiter/Rieder(2008): (Local) convergence for discretized setting.
- ▶ D-bar method: convergent 2D-implementation for $\sigma \in C^2$
Knudsen, Lassas, Mueller, Siltanen (2008)

In practice:

- ▶ large jumps in conductivity
- ▶ large interest in detecting shapes / inclusions / anomalies

Inclusion/shape detection problem:

$$\Lambda(\sigma) \mapsto \text{supp}(\sigma - \sigma_0)?, \quad \sigma_0: \text{reference conductivity.}$$

Shape detection

Promising approach: Factorization method (*Kirsch 1998*)

- ▶ *FM for EIT (1999–): Brühl, Hakula, Hanke, H., Hyvönen, Kirsch, Lechleiter, Nachman, Päivärinta, Pursiainen, Schappel, Schmitt, Seo, Teirilä*

Typical result:

$$z \notin \text{supp}(\sigma - \sigma_0) \quad \text{iff} \quad \lim_{\alpha \rightarrow 0} I_\alpha(z) = \infty.$$

($I_\alpha(z)$: *indicator function*)

Unsolved problems since 1998:

- ▶ Convergent regularization strategies for "infinity test"?
- ▶ Theory needs definiteness assumption, e.g., $\sigma \geq \sigma_0$ everywhere

In this talk: A monotonicity based sampling method.

$$\int_{\Omega} (\sigma_1 - \sigma_2) |\nabla u_1|^2 \, dx \leq (g, (\Lambda(\sigma_2) - \Lambda(\sigma_1))g)$$

u_1 solution corresponding to σ_1 and boundary current g .

Simple consequence:

$$\sigma_1 \geq \sigma_2 \implies \Lambda(\sigma_1) \leq \Lambda(\sigma_2)$$

Monotony based imaging

- ▶ True conductivity: $\sigma = 1 + \chi_D$, D : unknown inclusion
- ↪ $\Lambda(\sigma)$: measured data
- ▶ Test conductivity: $1 + \chi_B$, B : small ball
- ↪ $\Lambda(1 + \chi_B)$ can be simulated for different balls B

Monotony:

$$B \subseteq D \implies 1 + \chi_B \leq 1 + \chi_D = \sigma \implies \Lambda(1 + \chi_B) \geq \Lambda(\sigma)$$

Monotony based reconstruction algo. for EIT (*Tamburrino/Rubinacci 02*)

- ▶ For all B , calculate $\Lambda(1 + \chi_B)$ & test whether $\Lambda(1 + \chi_B) \geq \Lambda(\sigma)$
 - ↪ Result: upper bound of D .
-

Only an upper bound? Converse monotony relation?

Converse montony relation

Theorem (H./Ullrich)

$\Omega \setminus \overline{D}$ connected. $\sigma = 1 + \chi_D$.

$$B \not\subseteq D \implies \Lambda(1 + \chi_B) \not\cong \Lambda(\sigma).$$

\rightsquigarrow Monotony method detects exact shape.


(Extensions possible for non-connected complement, inhomogeneous inclusions or background, continuous transitions between inclusion and background, . . .)

Converse montony relation

Proof $(\sigma = 1 + \chi_D, \kappa = 1 + \chi_B)$

$$\int_{\Omega} (\kappa - \sigma) |\nabla u_{\kappa}|^2 dx \leq (g, (\Lambda(\sigma) - \Lambda(\kappa))g)$$

Apply localized potentials (H 2008) to control power term $|\nabla u_{\kappa}|^2$.



$$\rightsquigarrow \exists g : (g, (\Lambda(\sigma) - \Lambda(\kappa))g) \geq 0 \implies \Lambda(\sigma) \not\leq \Lambda(\kappa)$$

- ▶ Testing $\Lambda(1 + \chi_B) \geq \Lambda(\sigma)$ is expensive. One forw. prob. per B .

Theorem (H./Seo, SIAM J. Math. Anal. 2010)

Let κ, σ, σ_0 piecewise analytic and $\Lambda'(\sigma_0)\kappa = \Lambda(\sigma) - \Lambda(\sigma_0)$. Then

$$\text{supp}_{\partial\Omega}\kappa = \text{supp}_{\partial\Omega}(\sigma - \sigma_0).$$

$\text{supp}_{\partial\Omega}$: outer support (= support, if support is compact and has conn. complement)

↪ Replacing $\Lambda(1 + \chi_B)$ by its linear approx. should still recover the exact shape (linearization error does not affect the shape!)

Theorem (H./Ullrich)

Let $\Omega \setminus \overline{D}$ connected, $0 < k \leq 1/2$.

$$B \subseteq D \iff \Lambda(\mathbb{1}) + k\Lambda'(\mathbb{1})\chi_B \geq \Lambda(\sigma).$$

↪ Fast, requires only homogeneous forward solution

- ▶ Comp. cost equivalent to linearized methods or FM

(Again, extensions possible for non-connected complement, inhomogeneous inclusions or background, continuous transitions between inclusion and background, . . .)



Indefinite inclusions

Indefinite inclusions: $\sigma = 2 + \chi_{D^+} - \chi_{D^-}$.



With $\kappa^+ = 2 + \chi_B$. $\kappa^- = 2 - \chi_B$:

$$D^+ \cup D^- \subseteq B \iff \Lambda(\kappa^-) \geq \Lambda(\sigma) \geq \Lambda(\kappa^+).$$



- ▶ Indefinite inclusions can be treated by step-wise shrinking of larger test domains.
- ▶ Result can be linearized. Linearized version yields exact shape (no linearization error!)
- ▶ Result extends to general pcw. anal. conductivities (with some technical effort)



Reconstructions with exact data and with 0.1% noise.




Reconstructions with exact data and with 0.1% noise.



Reconstructions for smooth transitions between inclusion and background and for the indefinite case.

Goal: Enhance linearized methods



Standard linearized method vs. heuristic combination with FM
for frequency-difference EIT without ref. measurements

(H./Seo/Woo, IEEE Trans. Med. Imaging 2010)

New monotony based shape reconstruction method

- ▶ yields the exact shape not just an upper bound
- ▶ can be efficiently implemented by linearization (while still reconstructing the exact shape)

Advantages

- ▶ Rigorous treatment of indefinite inclusions
- ▶ Convergent regularization implementation of testing criteria seems possible

For practical applications:

- ▶ Enhance linearized/iterative methods by exact shape reconstruction (*H./Seo SIMA 2010, H./Seo/Woo IEEE TMI 2010*)